

Manual Supplement

Manual Title: 279 FC Users
Print Date: March 2016
Revision/Date: 1, 5/18

Supplement Issue: **2**
Issue Date: 4/22
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 697

On page 44, replace the **Temperature** section, under **Infrared Camera**:

Infrared Camera

Temperature

Temperature Measurement Range -10 °C to +200 °C

Temperature Measurement Accuracy ± 5 °C or ± 5 %, whichever is greater, at 25 °C (Ambient)

Temperature Coefficient 0.2 °C or 0.2 %, whichever is greater, for each °C from 25 °C
for target temperatures below 20 °C, add 0.2 °C for each degree °C

Change #2, 672

On page 44, replace the **Emissivity** under **Infrared Camera** with:

Infrared Camera

Emissivity 0.97 fixed